

Cell-Chip

Innovation by



The counting chamber looks like the familiar Neubauer „improved“ hemocytometer: The cells are distributed over 3 x 3 large squares, each with 1 mm edge length and with a surface area of 1 mm².

Count your cells as usual - With the Cell-Chip, you inject the sample, stained or unstained, into the desired chamber. Two separate counting chambers enable two counts per Cell-Chip.

Quick, easy and safe:

- Minimal counting tolerances
- High precision
- minimized risk of infection
- easy to recycle
- sterile, single wrapped

Product	Cat. No.	Dimensions	Volume	Depth of chamber	Pieces/ sterile unit	Pieces/ Box
Cell-Chip with counting grid Neubauer „improved“ Individually packaged	505050	25x75x1.6 mm	10 µl	0.1 mm	1 Chip (for 2 counts)	50 Chips

Useful products by Seraglob

Our label Seraglob provides scientists all over the world with first class serum, medium, reagents and additives.
Find more products and services at seraglob.com



Trypan Blue

Stain your cells with a colorant to facilitate counting

Cat. No. L 2001
Volume 100 ml
More at seraglob.com/additives



Fetal Bovine / Calf Serum


High quality serum to give your cells a head start

Cat. No. S 40500
Volume 500 ml
More at seraglob.com/sera

Details & Instructions

Structure of the „improved“ counting chamber

The counting chamber consists of 9 large squares (3x3), of which 4 are corner squares (L). The corner squares (L) are divided into 16 squares (4x4). The central square is divided into 5x5 squares (E) that are divided into 4x4.



Volume details for the L-squares

The area of the L-squares results from the edge lengths:

$$1 \text{ mm} \times 1 \text{ mm} = 1 \text{ mm}^2.$$

At a chamber depth of 0.1 mm this results in a volume of 0.1 mm^3 in the L-squares (conversion: 0.1 mm^3 correspond to $0.1 \mu\text{l}$ or 10^{-4} ml .)

Counting with the Cell-Chip

Leukocyte counting (1:20 dilution) <ol style="list-style-type: none"> 1. Dilute blood using accepted laboratory methods 2. Load $10 \mu\text{l}$ of diluted sample into the sample injection area 3. Count the erythrocytes in the 5 small squares (four small corner squares and one small middle square) of the large center square 	Amount of Leukocytes leukocytes per ml = (cells in 4 corner squares/ 4) x 20 (dilution factor) x 10^4 (volume factor)
Mammalian Cell counting <ol style="list-style-type: none"> 1. Treat the cell samples with Trypsin-EDTA. 2. Carefully remove the supernatant with a pipette tip without disturbing the pellet 3. Add an appropriate volume of growth media or PBS to dilute to a final concentration of 5×10^3 cells/ml to 5×10^6 cells per ml 4. Thoroughly resuspend the cell pellet with a pipette 5. Check visually if there are any cell clumps or agglomerates 6. Load $10 \mu\text{l}$ of sample into the sample injection area 7. Count the cells in 5 large squares 	Amount of Mammalian Cells mammalian cells per ml = (cells in 5 large squares/5) x dilution factor x 10^4 (volume factor)
Erythrocyte counting (1:200 dilution) <ol style="list-style-type: none"> 1. Dilute blood using accepted laboratory methods 2. Load $10 \mu\text{l}$ of diluted sample into the sample injection area 3. Count the erythrocytes in the 5 small squares (four small corner squares and one small middle square) of the large center square 	Amount of Erythrocytes erythrocytes per ml = cells in 5 small squares x 5 x 200 (dilution factor) x 10^4 (volume factor)